



SPARE PARTS RECOMMENDATIONS FOR MODEL 275HD

Level A, Consumable Parts:		
Part No.	Description	2 Yr. Req.
4704-1042	Screw 10-32 x 1-5/8" Pan Head Phillips 304 SS	
4903-0001	Gasket Ring For 4" 150# Flange Garlock 9850 1/16" Thk	
4903-0003	Gasket Ring For 3" 150# Flange	
4903-0004	Gasket Ring For 2" 150# Flange High Temp	
4903-1000	Gasket, Silicon Foam Closed Cell HT-800 1/4" Thick for 275 Enclosure	
4903-1001	Gasket, 275 Subflange	
4904-0016	O-Ring 2-217 Viton 75 Brown	1
4904-0036	O-Ring 2-318 Viton	1
4904-0041	O-Ring 2-012 Viton	6
4904-0046	O-Ring 2-110 Viton	2
4904-1006	O-Ring 2-017 Viton	4
4955-0136	Valve SV 120VAC 2Way High Flow	
4955-0137	Valve SV 2Way 24Vdc High Flow Parker	
4955-1002	Valve, Check 1 Psig Cracking Pressure 1/4" Tube Fittings Both Ends	
4956-0001	Accumulator Sphere 7" SS	
4980-0327	Boss Mount, .437" Safety Screen 0.004" Hole, 0.09Sqin Open Area	1
4980-0328	Filter Element 316SS 9" 2U OD 30mm X 235mm L X 2 Micron	2

Eductor (Sample Dilution) Critical Orifice Options		
Part No.	Description	2 Yr. Req.
4990-1023	Orifice #3 for Critical Orifice with Orifice holding bolt	
4990-1024	Orifice #4 for Critical Orifice with Orifice holding bolt	
4990-1025	Orifice #5 for Critical Orifice with Orifice holding bolt	
4990-1026	Orifice #6 for Critical Orifice with Orifice holding bolt	
4990-1027	Orifice #7 for Critical Orifice with Orifice holding bolt	
4990-1028	Orifice #8 for Critical Orifice with Orifice holding bolt	
4990-1029	Orifice #9 for Critical Orifice with Orifice holding bolt	
4990-1030	Orifice #10 for Critical Orifice with Orifice holding bolt	
4990-1031	Orifice #11 for Critical Orifice with Orifice holding bolt	
4990-1032	Orifice #12 for Critical Orifice with Orifice holding bolt	
4990-1033	Orifice #15 for Critical Orifice with Orifice holding bolt	
5110-0015	Acid Blocker Assembly	
5209-0380	Dilution Critical Orifice #3 Sub Assembly	
5209-0382	Dilution Critical Orifice #5 Sub Assembly	
5209-0383	Dilution Critical Orifice #6 Sub Assembly	
5209-0384	Dilution Critical Orifice #7 Sub Assembly	
5209-0385	Dilution Critical Orifice #8 Sub Assembly	
5209-0386	Dilution Critical Orifice #9 Sub Assembly	
5209-0387	Dilution Critical Orifice #10 Sub Assembly	
5209-0388	Dilution Critical Orifice #11 Sub Assembly	
5209-0389	Dilution Critical Orifice #12 Sub Assembly	
5209-0390	Dilution Critical Orifice #15 Sub Assembly	
5209-1000	Sub Assembly Fast-Loop Orifice Bolt for 275HD	
5209-1002	Manifold, 275HD, SS, w/ Fittings, Use w/o HPA ***Does Not Include O-Rings***	
5209-1005	Assembly Filter and Cap for 275HD Chamber w/ Woven Filter SS	

Continued on next page

UNIVERSAL ANALYZERS INC.
 5200 Convair Drive Carson City, NV 89706, USA
 Telephone: +1 (775) 883-2500
 Fax: +1 (775) 883-6388
 www.UniversalAnalyzers.com



Optional Parts:		
Part No.	Description	2 Yr. Req.
8010-0011	Silicone Grease 1 Ounce 700F High Temp For Kits	
8000-0016	Tape High Temperature Poly-Temp Nickel Anti-Seize (Specs)	
8010-1000	Capsule Vapor 2' Two Year Corrosion Protection	
9506-1001	Cleaning Wire for Eductor and #15 Orifice 12 mil	
9506-1002	Brush, Flue Double Spiral High Carbon Steel	
9506-1005	Flue Brush Extension Handle 36"	
9506-1006	T Extentsion Flue Brush Handle	
9515-1000	Kit Lubricant 2450 Deg F Never Seez Heavy Metal Free for Severe Industrial Applications	
3014-1001	Heater Element 3/8" OD x 11" long w/ 24" total lead length 800W 240 VAC .	
1150-1002	Sensor RTD PT100 2" Sheath 3/16" Diameter 1/8" MNPT Mount with 60" Armor Leads	
1150-1001	Sensor Thermocouple Type K 2" Sheath 3/16" Diameter 1/8" Mnpt Mount with 60" Armor Leads	